




G2938-80001



Certificate of Conformity to Manufacturing Specification

Agilent part number:	G2938-80001	Lot number:	2333
Product name:	Agilent Small RNA Reagents Part II	Kit expiration date:	14-Aug-2024
Included in:	5067-1548, 5067-1549 and 5067-1550		
Storage conditions:	-28 °C to -13 °C	Approved by:	Jurgita Zilinskiene
Shipping conditions:	2 °C to 8 °C	Date:	30-Aug-2023
			Thermo Fisher Scientific Baltics UAB Vilnius, Lithuania

This product was manufactured and verified in accordance with an ISO 9001 certified quality assurance system. The manufacture, inspection, and testing have been performed by trained personnel. The instrumentation and materials used in the qualification process are identified and traceable. All associated test instruments are maintained and calibrated.

The following characteristics have been tested for conformity to manufacturing specifications.

Product characteristic	Method	Result
Small RNA Ladder		
Concentration	Photometrical	Pass
Lot Consistency	Electrophoretical	Pass
Stability	Electrophoretical	Pass

For Research Use Only. Not for use in diagnostic procedures.

Agilent Technologies
Deutschland GmbH
Hewlett-Packard-Str. 8
76337 Waldbronn, Germany

Made in Lithuania

For global support: www.agilent.com/en/contact-us/page

Template Number: G2938-80001 CoC
Template Revision: A.00