




G2938-80021



Certificate of Conformity to Manufacturing Specification

| | |
|--|---|
| Agilent part number: G2938-80021 | Lot number: 2541 |
| Product name: Agilent DNA 7500 Reagents | Kit expiration date: 06-Oct-2026 |
| Included in: 5067-1506 and 5067-1507 |  |
| Storage conditions: 2 °C to 8 °C | Approved by: Jurgita Zilinskiene |
| Shipping conditions: 2 °C to 8 °C | Date: 15-Oct-2025 |
| | Thermo Fisher Scientific Baltics UAB Vilnius, Lithuania |

This product was manufactured and verified in accordance with an ISO 9001 certified quality assurance system. The manufacture, inspection, and testing have been performed by trained personnel. The instrumentation and materials used in the qualification process are identified and traceable. All associated test instruments are maintained and calibrated.

The following characteristics have been tested for conformity to manufacturing specifications.

| Product characteristic | Method | Result |
|----------------------------|-------------------|--------|
| DNA 7500 Gel Matrix | | |
| Conductivity | Electrochemical | Pass |
| pH | Electrochemical | Pass |
| No DNase contamination | Electrophoretical | Pass |
| Resolution | Electrophoretical | Pass |
| DNA 7500 Markers | | |
| Conductivity | Electrochemical | Pass |
| pH | Electrochemical | Pass |
| Marker | Electrophoretical | Pass |
| No DNase contamination | Electrophoretical | Pass |
| DNA 7500 Ladder | | |
| Concentration | Photometrical | Pass |
| Lot Consistency | Electrophoretical | Pass |
| No DNase contamination | Electrophoretical | Pass |
| DNA Dye Concentrate | | |
| Concentration | Photometrical | Pass |

For Research Use Only. Not for use in diagnostic procedures.

Agilent Technologies
Deutschland GmbH
Hewlett-Packard-Str. 8
76337 Waldbronn, Germany

Made in Lithuania

For global support: www.agilent.com/en/contact-us/page

Template Number: G2938-80021 CoC
Template Revision: A.00