




G2938-80022



Certificate of Conformity to Manufacturing Specification

Agilent part number: G2938-80022	Lot number: 2422
Product name: Agilent DNA 12000 Reagents	Kit expiration date: 27-May-2025
Included in: 5067-1508 and 5067-1509	
Storage conditions: 2 °C to 8 °C	Approved by: Jurgita Zilinskiene
Shipping conditions: 2 °C to 8 °C	Date: 05-Jun-2024
	Thermo Fisher Scientific Baltics UAB Vilnius, Lithuania

This product was manufactured and verified in accordance with an ISO 9001 certified quality assurance system. The manufacture, inspection, and testing have been performed by trained personnel. The instrumentation and materials used in the qualification process are identified and traceable. All associated test instruments are maintained and calibrated.

The following characteristics have been tested for conformity to manufacturing specifications.

Product characteristic	Method	Result
DNA 12000 Gel Matrix		
Conductivity	Electrochemical	Pass
pH	Electrochemical	Pass
No DNase contamination	Electrophoretical	Pass
Resolution	Electrophoretical	Pass
DNA 12000 Markers		
Conductivity	Electrochemical	Pass
pH	Electrochemical	Pass
Marker	Electrophoretical	Pass
No DNase contamination	Electrophoretical	Pass
DNA 12000 Ladder		
Concentration	Photometrical	Pass
Lot Consistency	Electrophoretical	Pass
No DNase contamination	Electrophoretical	Pass
DNA Dye Concentrate		
Concentration	Photometrical	Pass

For Research Use Only. Not for use in diagnostic procedures.