




G2938-80025



Certificate of Conformity to Manufacturing Specification

| | | | |
|-----------------------------|------------------------------------|---|--|
| Agilent part number: | G2938-80025 | Lot number: | 2424 |
| Product name: | Agilent Protein 80 Reagents Part I | Kit expiration date: | 10-Mar-2025 |
| Included in: | 5067-1515 and 5067-1516 |  | |
| Storage conditions: | 2 °C to 8 °C | Approved by: | Jurgita Zilinskiene |
| Shipping conditions: | 2 °C to 8 °C | Date: | 21-Jun-2024 |
| | | | Thermo Fisher Scientific Baltics UAB Vilnius, Lithuania |

This product was manufactured and verified in accordance with an ISO 9001 certified quality assurance system. The manufacture, inspection, and testing have been performed by trained personnel. The instrumentation and materials used in the qualification process are identified and traceable. All associated test instruments are maintained and calibrated.

The following characteristics have been tested for conformity to manufacturing specifications.

| Product characteristic | Method | Result |
|--------------------------------|-----------------|--------|
| Protein 80 Gel Matrix | | |
| Conductivity | Electrochemical | Pass |
| pH | Electrochemical | Pass |
| Protein Dye Concentrate | | |
| Concentration | Photometrical | Pass |

For Research Use Only. Not for use in diagnostic procedures.