G2938-80026

# Certificate of Conformity to Manufacturing Specification

**Agilent part number:** G2938-80026

**Product name:** Agilent Protein 230 Reagents Part I

**Included in:** 5067-1517 and 5067-1518

**Storage conditions:** 2 °C to 8 °C

**Shipping conditions:** 2 °C to 8 °C

**Lot number:** 2638

**Kit expiration date:** 14-Jun-2027

**Approved by:** Sebastian Pallmann
Director Quality Science & Technology (QSAT)
Agilent Technologies

**Date:** 21-Sep-2026

This product was manufactured and verified in accordance with an ISO 9001 certified quality assurance system. The manufacture, inspection, and testing have been performed by trained personnel. The instrumentation and materials used in the qualification process are identified and traceable. All associated test instruments are maintained and calibrated.

The following characteristics have been tested for conformity to manufacturing specifications.

| Product characteristic | Method | Result |
|---|---|---|
| **Protein 230 Gel Matrix** | | |
| Conductivity | Electrochemical | Pass |
| pH | Electrochemical | Pass |
| **Protein Dye Concentrate** | | |
| Concentration | Photometrical | Pass |

**For Research Use Only. Not for use in diagnostic procedures.**

Agilent Technologies
Deutschland GmbH
Hewlett-Packard-Str. 8
76337 Waldbronn, Germany

**Made in Lithuania**

**For global support: www.agilent.com/en/contact-us/page**

Template Number: G2938-80026 CoC
Template Revision: A.01