




G2938-85003



## Certificate of Conformity to Manufacturing Specification

<b>Agilent part number:</b>	G2938-85003	<b>Lot number:</b>	2515
<b>Product name:</b>	Agilent High Sensitivity Protein 250 Ladder	<b>Kit expiration date:</b>	07-Apr-2026
<b>Included in:</b>	5067-1575 and 5067-1578	<b>Approved by:</b>	 Jurgita Zilinskiene
<b>Storage conditions:</b>	-28 °C to -13 °C	<b>Date:</b>	17-Apr-2025
<b>Shipping conditions:</b>	below -13 °C		Thermo Fisher Scientific Baltics UAB Vilnius, Lithuania

This product was manufactured and verified in accordance with an ISO 9001 certified quality assurance system. The manufacture, inspection, and testing have been performed by trained personnel. The instrumentation and materials used in the qualification process are identified and traceable. All associated test instruments are maintained and calibrated.

The following characteristics have been tested for conformity to manufacturing specifications.

Product characteristic	Method	Result
<b>High Sensitivity Protein 250 Ladder</b>		
Concentration	Photometrical	Pass
Functional	Electrophoretical	Pass

For Research Use Only. Not for use in diagnostic procedures.

Agilent Technologies  
Deutschland GmbH  
Hewlett-Packard-Str. 8  
76337 Waldbronn, Germany

Made in Lithuania

For global support: [www.agilent.com/en/contact-us/page](http://www.agilent.com/en/contact-us/page)

Template Number: G2938-85003 CoC  
Template Revision: A.00