




G2938-85004



Certificate of Conformity to Manufacturing Specification

| | |
|--|---|
| Agilent part number: G2938-85004 | Lot number: 2402 |
| Product name: Agilent High Sensitivity DNA Reagents | Kit expiration date: 08-Jan-2025 |
| Included in: 5067-4626 and 5067-4627 |  |
| Storage conditions: 2 °C to 8 °C | Approved by: Jurgita Zilinskiene |
| Shipping conditions: 2 °C to 8 °C | Date: 19-Jan-2024 |
| | Thermo Fisher Scientific Baltics UAB Vilnius, Lithuania |

This product was manufactured and verified in accordance with an ISO 9001 certified quality assurance system. The manufacture, inspection, and testing have been performed by trained personnel. The instrumentation and materials used in the qualification process are identified and traceable. All associated test instruments are maintained and calibrated.

The following characteristics have been tested for conformity to manufacturing specifications.

| Product characteristic | Method | Result |
|--|-------------------|--------|
| High Sensitivity DNA Gel Matrix | | |
| Conductivity | Electrochemical | Pass |
| pH | Electrochemical | Pass |
| No DNase contamination | Electrophoretical | Pass |
| High Sensitivity DNA Markers | | |
| Conductivity | Electrochemical | Pass |
| pH | Electrochemical | Pass |
| Marker | Electrophoretical | Pass |
| No DNase contamination | Electrophoretical | Pass |
| High Sensitivity DNA Ladder | | |
| Functional | Electrophoretical | Pass |
| Lot Consistency | Electrophoretical | Pass |
| No DNase contamination | Electrophoretical | Pass |
| DNA Dye Concentrate | | |
| Concentration | Photometrical | Pass |

For Research Use Only. Not for use in diagnostic procedures.