



CERTIFICATE OF ANALYSIS

Agilent Product Name: Reference Solution

Agilent Part No: 5190-0545-1 in Kit 5190-0545/5190-3988/5190-3990/5190-3992/5190-3999
5190-4003/5190-4002

Lot No.: 14330001

Quality Statement

This electrode solution was manufactured and verified in accordance with The Metrological Law of the People's Republic of China and an ISO 9001 System.

Quality Specifications

Component	Concentration (mg/mL)
Potassium Chloride	233 ± 5%
Silver Chloride	3 ± 5%

Manufacturing Statement:

Purified reagents and deionized water were used in manufacture. Glassware used in manufacture was calibrated in accordance with GB12806 and JJG196. Balances used in manufacture are calibrated in accordance with JB5374, JJG1036 and GB/T26497.

Storage:

Store at room temperature (18 – 25°C).
Extended storage at temperatures below 4°C or above 35°C is not recommended.

Certified by:

Neil Noronha
Chemistries and Supplies Division
Quality Manager

Date of expiration: 12 Nov. 2015