


Paraquat in Water



Sample: Paraquat and Diquat, 10 μ L/injection
Column: Hypersil ODS, 2.1 x 100 mm, 5 μ m, (P/N: 799160D - 552)
Mobile phase: 0.1% hexanesulfonic acid
0.35% triethylamine pH 2.5 (H_3PO_4)
Flow rate: 0.4 mL/min
Detection: 256/10 nm and 310/10 nm (reference 450/100 nm)