

1-kb DNA Standards Ladder

Column: μ SIL-DNA
75 μ m I.D. x 65 cm (20 cm to detector)

J&W P/N: 199-2602

Buffer: 0.5% HPMC/1x TBE, pH 8.3

Voltage: -160 V/cm

Injector: electrokinetic, -5 kV for 3 sec

Temperature: ambient

Detector: UV, 260 nm

Sample: 0.2 μ g/ μ L of 1-kb DNA standards ladder
Membrane dialysis for 30 min at 4°C before injection

