


## Supercoiled DNA Ladder

Column:  $\mu$ SIL-FC  
50  $\mu$ m I.D. x 60 cm (20 cm to detector)

J&W P/N: 194-8111  
Buffer: 0.1% HPMC/1x TBE, pH 8.3  
Voltage: -6 kV  
Injector: electrokinetic, -7.5 kV for 5 sec  
Temperature: ambient  
Detector: UV, 260 nm  
Sample: Supercoiled DNA ladder (Courtesy of Life Technologies)  
Membrane dialysis for 30 min at 4°C before injection



CE021