

Base Neutral 3

Column: DB-5.625

30 m x 0.25 mm I.D., 0.25 µm

J&W P/N: 122-5631

Carrier: Helium at 33 cm/sec


Oven: 35°C for 5 min
35-280°C at 8°/min
280°C for 20 min

Injector: Splitless, 250°C

Detector: FID, 300°C
Nitrogen makeup
gas at 30 mL/min

Compounds and Retention Times:

- 28. Hexachloroethane (15.13)
- 29. N-Nitrosodi-n-propylamine (15.13)
- 30. Isophorone (16.23)
- 31. 1,2,4-Trichlorobenzene (17.41)
- 32. Hexachlorocyclopentadiene (20.53)
- 33. 2-Chloronaphthalene (21.29)
- 34. N-Nitrosodiphenylamine (25.40)
- 35. Azobenzene (25.40)
- 36. Phenanthrene (27.83)
- 37. Fluoranthene (31.63)
- 38. Benzylbutylphthalate (34.77)



C020