

Amines in Water

Column: CAM
30 m x 0.25 mm I.D., 0.25 μ m
J&W P/N: 112-2132
Carrier: Hydrogen at 38 cm/sec
Oven: 120-220°C at 10°/min
Injector: Split
Detector: FID
Nitrogen makeup gas at 30 mL/min

