

Primary Amines

Column: CAM
30 m x 0.25 mm I.D., 0.25 μ m
J&W P/N: 112-2132
Carrier: Hydrogen at 40 cm/sec
Oven: 110°C isothermal
Injector: Split
Detector: FID
Nitrogen makeup gas at 30 mL/min

1. *n*-Octylamine
2. *n*-Nonylamine
3. *n*-Decylamine
4. Benzylamine
5. Dicyclohexylamine

