


Simulated Distillation

Column: DB-2887
10 m x 0.53 mm I.D., 3.0 µm
J&W P/N: 125-2814
Carrier: Helium at 7 mL/min
Oven: 35-350°C at 15°/min
Injector: Megabore Direct
Detector: FID
Nitrogen makeup gas
at 30 mL/min

1. *n*-Pentane
2. *n*-Hexane
3. *n*-Heptane
4. *n*-Octane

5. *n*-Nonane
6. *n*-Decane
7. *n*-Undecane
8. *n*-Dodecane
9. *n*-Tetradecane
10. *n*-Pentadecane
11. *n*-Hexadecane
12. *n*-Heptadecane
13. *n*-Octadecane
14. *n*-Eicosane
15. *n*-Tetracosane
16. *n*-Octacosane
17. *n*-Dotriacontane
18. *n*-Hexatriacontane
19. *n*-Tetracontane



C166