

EPA Method 551

Column: DB-210
30 m x 0.25 mm I.D., 0.5 µm
J&W P/N: 122-0233
Carrier: Helium at 24.8 cm/sec, measured at 150°C
Oven: 35°C for 11 min
35-120°C at 10°/min
120°C for 19 min
Injector: Splitless, 200°C
15 sec purge activation time
1 µL of 50 pg/µL, AccuStandard
Detector: ECD, 300°C

1. Chloroform
2. 1,1,1-Trichloroethane
3. Carbon tetrachloride
4. Trichloroacetonitrile
5. Dichloroacetonitrile
6. Bromodichloromethane
7. Trichloroethylene
8. Chloral hydrate
9. 1,1-Dichloro-2-propanone
10. Dibromochloromethane
11. Bromochloroacetonitrile
12. 1,2-Dibromoethane
13. Tetrachloroethylene
14. 1,1,1-Trichloropropanone
15. 1,2-Dibromopropane (IS)
16. Bromoform
17. Dibromoacetonitrile
18. 1,2-Dibromo-3-chloropropane

