

Volatiles (EPA Method 1311)

Column: DB-5.625
30 m x 0.25 mm I.D., 0.5 µm
J&W P/N: 122-5632
Carrier: Helium at 35 cm/sec, measured at 40°C
Oven: 40°C for 1 min
40-140°C at 12°/min
Injector: Split 1:40, 250°C
1 µL of 0.5-1.0 µg/µL standard in methanol
Detector: MSD, 280°C transfer line
full scan at m/z 35-200

- | | |
|-----------------------|-------------------------|
| 1. 1,1-Dichloroethene | 7. Carbon tetrachloride |
| 2. Methylene chloride | 8. Trichloroethene |
| 3. 2-Butanone (MEK) | 9. Pyridine |
| 4. Chloroform | 10. Tetrachloroethene |
| 5. 1,2-Dichloroethane | 11. Chlorobenzene |
| 6. Benzene | |

