

Butter Triglycerides

Column: DB-17ht
30 m x 0.32 mm I.D., 0.15 μ m
J&W P/N: 123-1831
Carrier: Hydrogen at 40 cm/sec
Oven: 250-365°C at 5°/min
365°C for 1 min
Injector: Cool On-column
1 μ L of 9 μ g/ μ L in toluene
(approx. 1% w/w solution)
Detector: FID, 400°C
Nitrogen makeup gas at 30 mL/min
Baseline Corrected

