

Triethylene Glycol and Impurities

Column: DB-1
30 m x 0.45 mm I.D., 1.27 μ m
J&W P/N: 124-1032
Carrier: Helium at 35 cm/sec,
measured at 50°C
Oven: 170°C isothermal
Injector: Split 1:50, 0.5 μ L, 250°C
Detector: FID, 280°C
Nitrogen makeup gas at 30 mL/min

