


Hallucinogens

Column: **DB-17ms**
30 m x 0.25 mm I.D., 0.25 µm
J&W P/N: 122-4732

Carrier: Helium at 30 cm/sec, measured at 50°C
Oven: 50°C for 0.5 min
125-255°C at 25°/min
255-320°C at 25°/min
320°C for 16 min
Injector: Splitless, 250°C
30 sec purge activation time
1 µL of 10-50 ng/µL standard in methanol
Detector: MSD, 300°C transfer line
full scan at m/z 40-350

1. 4-Methylaminorex
2. Mescaline
3. Dimethyltryptamine
4. Ketimine
5. TCP (Tenocyclidine)
6. PCP
7. Diethyltryptamine
8. Bufotenine
9. Fentanyl
10. Ibogaine
11. LSD
12. Psilocin (peak not shown)



C468