

Agilent ULTRON ES

A collection of citations to advance your research

Table of contents

Food testing and agriculture

Food testing and agriculture

Misleading measures in Vitamin D analysis: A novel LC-MS/MS assay to account for epimers and isobars

Nutrition Journal, **10** (2011) Iltaf Shah *et al.*

Tags

Bond Elut SI, Bond Elut Plexa, Bond Elut LMS, Bond Elut PPL, SampliQ OPT, SampliQ DVB, ZORBAX RRHD SB-C18, Ultron ES-OVM, food testing & agriculture, dietary supplements, natural compounds & additives

Abstract

Recently, the accuracies of many commercially available immunoassays for Vitamin D have been questioned. Liquid chromatography tandem mass spectrometry (LC- MS/MS) has been shown to facilitate accurate separation and quantification of the major circulating metabolite 25-hydroxyvitamin-D3 (250HD3) and 25-hydroxyvitamin-D2 (250HD2) collectively termed as 250HD. However, among other interferents, this method may be compromised by overlapping peaks and identical masses of epimers and isobars, resulting in inaccuracies in circulating 250HD measurements. The aim of this study was to develop a novel LC-MS/MS method that can accurately identify and quantitate 250HD3 and 250HD2 through chromatographic separation of 250HD from its epimers and isobars. © The Authors.

www.agilent.com/chem

Agilent shall not be liable for errors contained herein or for incidental or consequential damages in connection with the furnishing, performance, or use of this material. Information, descriptions, and specifications in this publication are subject to change without notice.

© Agilent Technologies, Inc., 2013

Printed in the UK October 1, 2013

5991-3048EN



