



**Agilent Technologies**

**Agilent J&W CP-Sil 8 CB for Amines**

A collection of citations to advance your research

**Table of contents**

[Energy and chemicals](#)

# Energy and chemicals

## [Capillary GC column performance test for basic analytes](#)

*Journal of Microcolumn Separations*, **10**, 661–669 (1998)

R. J. M. N. Dresen, H. J. W. Henderickx, S. van der Wal

### Tags

CAM, CP Sil 8 CB for Amines, energy and chemicals

### Abstract

An investigation of GC columns for basic analytes found that Agilent J&W CP Sil 8 CB for Amines and CAM columns were best suited for analysis of amines at low concentration levels with respect to irreversible adsorption. Published by John Wiley and Sons.

---

[www.agilent.com/chem](http://www.agilent.com/chem)

Agilent shall not be liable for errors contained herein or for incidental or consequential damages in connection with the furnishing, performance, or use of this material. Information, descriptions, and specifications in this publication are subject to change without notice.

© Agilent Technologies, Inc., 2014

Printed in the UK  
March 10, 2014

5991-4213EN

The Measure of Confidence



**Agilent Technologies**