Agilent Sulfur Chemiluminescence Detector and Nitrogen Chemiluminescence Detector

Specification Guide

Agilent 8355 SCD

MDL[1]	Typical <0.5 pg(S)/sec
MDL (SCD/FID Tandem)	Typical <5.0 pg(S)/sec
Typical selectivity g S/g C[2]	$>2 \times 10^7$
Linearity[3]	$>10^4$
Repeatability	<2% RSD 2 hours <5% RSD 24 hours
Ozone supply gas	UHP oxygen, UHP Certified Grade Mixture 80% Ar/20% O_2
Oxidizer	Ultra Zero Grade air
Hydrogen	Zero Grade
Analog output	0–1 V, 0–10 V (Standalone version only)

[1] **MDL:** Burner temperature 800 °C, 8 mL/min lower hydrogen, 38 mL/m upper hydrogen, 50 mL/min air, *tert*-butyl disulfide in SCD checkout sample (5190-7003) as the test compound, 30 m × 0.32 mm, 1 µm DB1 (123-1033), 50 °C for 3 minutes, 25 °C/min to 160 °C, hold 2 minutes, 1 µL splitless injection, fully stabilized burner.

[2] **Selectivity:** Defined as the sensitivity of S over the sensitivity of a selected hydrocarbon. Operating parameters same as MDL. Isooctane (sample solvent) and *tert*-butyl disulfide in the SCD checkout sample are used as the testing compounds. Selectivity performed on a fully stabilized burner.

[3] **Linearity:** Operating parameters same as MDL except 12 mL/min lower hydrogen and 38 mL/m upper hydrogen; 80 °C for 1 minute, 25 °C/min to 160 °C, hold 2 minutes, COC inlet; Test compound: *tert*-butyl disulfide in isooctane. Linearity performed on a fully stabilized burner.

Agilent Technologies

Agilent 8255 NCD

MDL[1]	Typical <3 pg(N)/sec
MDL (NCD/FID Tandem)[1]	Typical <30 pg(N)/sec
Typical selectivity g N/g C[2]	$>2 \times 10^7$
Typical selectivity g N/g C (NCD/FID Tandem)	$>10^6$
Linearity[3]	$>10^4$
Repeatability	<1.5% RSD 8 hours <2% RSD 18 hours
Ozone supply gas	UHP oxygen
Oxidizer	UHP oxygen
Hydrogen	Zero Grade
Analog output	0–1 V, 0–10 V (Standalone version only)

1 **MDL:** Burner temperature 900 °C, 3 mL/min hydrogen, 8 mL/m oxygen, 3-methylindole in NCD checkout sample (5190-7002) as the test compound, 30 m × 0.32 mm, 0.25 µm HP-5 (19091J-413), 50 °C for 3 minutes, 25 °C/min to 250 °C, hold 2 minutes, 1 µL splitless injection, fully stabilized burner. NCD/FID configuration requires oxygen for the oxidizer and helium for the make-up gas.

2 **Selectivity:** Defined as the sensitivity of N over the sensitivity of a selected hydrocarbon. Operating parameters same as MDL. Isooctane (sample solvent) and 3-methylindole in the NCD checkout sample are used as the testing compounds. Selectivity performed on a fully stabilized burner.

3 **Linearity:** Operating parameters same as MDL except oven temperatures (80 °C for 1 minute, 25 °C/min to 180 °C, hold 1 minute, nitrobenzene in isooctane as test compound. Linearity performed on a fully stabilized burner. In the NCD/FID configuration, the FID linearity is 10^6.

Physical Specifications

Power requirements

8255/8355 Detector and Pump	120/220–240 V 50/60 Hz 1,200 VA

Dimensions

Detector	Height: 41.0 cm (16.1 in) Width: 27.0 cm (10.6 in) Depth: 51.1 cm (20.1 in)
8355 SCD weight	22 kg (49 lbs)
8255 NCD weight	24 kg (52 lbs)
Burner	Height: 22.1 cm (8.7 in) Weight: 0.7 kg (1.5 lbs)
Vacuum pump	Height: 26.1 cm (10.3 in) Width: 15.8 cm (6.2 in) Depth: 43.0 cm (16.9 in) Weight: 25 kg (55 lbs)

Environmental conditions

Installation category	II
Pollution degree	2
Ambient temperature	50–104 °F (10–40 °C)
Relative humidity	80% at 87.5 °F (31 °C) 50% at 104 °F (40 °C)
Normal operating environment	Intended for indoor use only
Maximum altitude	2,000 m (6,562 ft)

www.agilent.com

Agilent shall not be liable for errors contained herein or for incidental or consequential damages in connection with the furnishing, performance, or use of this material.

Information, descriptions, and specifications in this publication are subject to change without notice.

© Agilent Technologies, Inc., 2016–2018
Printed in the USA
February 2, 2018
5991-6379EN

Agilent Technologies