

6-Column and 8-Row Reservoir Plates

Part Number	201284-100	204284-100	201260-100	201282-100	204019-100	204367-100
Product Description	Reservoir, 6 column, polypropylene, 47 mL/column, 282 mL maximum, 24 pyramid base geometries, 44 mm height, 25/pk	Reservoir, 6 column, polypropylene, 47 mL/column, 282 mL maximum, 24 pyramid base geometries, 44 mm height, irradiated, 25/pk	Reservoir, 8 row, polypropylene, 32 mL/row, 256 mL maximum, pyramid base geometries, 44 mm height, 25/pk	Reservoir, 8 row, polypropylene, 10.6 mL/row, 85 mL maximum, pyramid base geometries, 19 mm height, 25/pk	Reservoir, 8 row, irradiated polypropylene, 32 mL/row, 256 mL maximum, pyramid base geometries, 44 mm height, 25/pk	Reservoir, 8 row, polypropylene, 10.6 mL/row, 85 mL max, pyramid base geometries, 19 mm height, irradiated, 25/pk
Specifications						
Well Number	6	6	8	8	8	8
Max Well Volume (mL)	47.28	47.28	32.50	10.90	32.50	10.90
Well Format	Open	Open	Open	Open	Open	Open
Bottom Shape	Pyramid column	Pyramid column	Pyramid row	Pyramid row	Pyramid row	Pyramid row
Bottom Shape Number	24	24	8	8	8	8
Dimensions (L × W) (mm)	127.76 × 85.47	127.76 × 85.47	127.76 × 85.47	127.76 × 85.47	127.76 × 85.47	127.76 × 85.47
Plate Height (mm)	44.04	44.04	44.04	19.05	44.04	19.05
Well Depth (mm)	42.08	42.08	39.22	12.98	39.22	12.98
Material	Polypropylene	Polypropylene	Polypropylene	Polypropylene	Polypropylene	Polypropylene
Color	Natural	Natural	Natural	Natural	Natural	Natural
Irradiated	No	Yes	No	No	Yes	Yes
Also Available as Irradiated	204284-100	N/A	204019-100	204367-100	N/A	N/A
Feature						
Shelf Life*	Non-irradiated parts: Best used within 5 years of production date.					
SHELLITE	Irradiated parts: Expire 5 years post production date.					
Packaging						
Plate/Case	25	25	25	25	25	25
Additional Information						
Microplate facility is a DNase/RNase free production environment with ISO 9001:2015 operations. All plates are desirand and approximate the production environment with the ANGLOGO X 2004 are siferable as						

- All plates are designed and manufactured in accordance with the ANSI/SBS X-2004 specifications.
- All reservoirs are designed to comply with ANSI/SLAS 1-2004: Microplates.
 Footprint dimensions are compatible with most automation systems.

www.agilent.com/lifesciences/microplates

DE-002183

This information is subject to change without notice.

© Agilent Technologies, Inc. 2022, 2024 Printed in the USA, October 24, 2024 5994-4428EN

^{*}Products should be stored in the original sealed package under normal laboratory environment conditions.