



Fiber Optic Dip Probe Coupler

Installation category I
Pollution degree 2
Equipment class III

Introduction

The Fiber Optic Dip Probe Coupler is used in conjunction with a fiber optic dip probe (ordered separately) to facilitate rapid sample measurement with your Cary spectrophotometer.

The instructions for installing the accessory are provided online with the Cary WinUV software. To view these instructions, start the Cary WinUV software, choose 'Help Topics' from the 'Help' menu and select 'Accessories' from the Cary WinUV Help home page. Click the 'Fiber Optic Dip Probe Coupler' option from the list of accessories for your instrument model.

Safety information

Before using this accessory, you must read the Safety Practices and Hazards section in your Cary operating instructions.

If the equipment is used in a manner not specified by the manufacturer, the protection provided by the equipment may be impaired.

Specifications

For environmental and power supply requirements, refer to your Cary operating instructions.

Connection

3.5 mm phono jack plugged into the port in the left side of the Cary sample compartment for the remote read switch.

Weights

Packed: 3.25 kg (7.2 lb)
Unpacked: 2.2 kg (4.8 lb)

Dimensions (W x D x H)

Packed: 370 x 360 x 220 mm (14.6 x 14.1 x 8.7 in)
Unpacked: 120 x 200 x 305 mm (4.7 x 7.9 x 12 in)



Agilent Technologies

This information is subject to change without notice.



8510159700

Part Number: 8510159700

Edition 04/12

Issue 5

© Agilent Technologies, Inc. 1997, 1999,
2011 and 2012

Agilent Technologies
679 Springvale Road
Mulgrave, VIC 3170