



400714 0006365362



**Agilent Technologies**

Lot 0006365362

Page 1 of 1

---

Catalog Number	400714
Product Name	StrataClean Resin
Expiration Date	2019-07-30
Quantity	3 ml
Certified By	Tim Sanchez
Quality Controlled By	Matt Huffman
Shipping Conditions	Shipped with an ice pack.
Storage Conditions	Store at 4°C.

**Test Conditions**

StrataClean resin is used to extract restriction enzymes, which are then incubated with 1 µg of the pBluescript II phagemid for 18 hours at the appropriate temperature. The samples are run on a 0.8% (w/v) agarose gel and checked for lack of cutting or nicking.

**Limited Product Warranty**

This warranty limits our liability to replacement of this product. No other warranties of any kind, express or implied, including, without limitation, implied warranties of merchantability or fitness for a particular purpose, are provided by Agilent. Agilent shall have no liability for any direct, indirect, consequential, or incidental damages arising out of the use, the results of use, or the inability to use this product.

For Research Use Only. Not for use in diagnostic procedures. This certificate is a declaration of analysis at the time of manufacture.

**Made in USA**

Agilent Technologies  
1834 State Highway 71 West  
Cedar Creek, TX 78612 USA

For support within the US or Canada: 1-800-227-9770-3-4  
For global support: [www.agilent.com/genomics/contactus](http://www.agilent.com/genomics/contactus)