



5185-5988 101564466



Lot 101564466

Page 1 of 1

---

Catalog Number	5185-5988
Product Name	Buffer B, Elution
Expiration Date	2025-04-30
Quantity	1 L
Certified By	Todd Parsons
Quality Controlled By	Matt Huffman
Shipping Conditions	Shipped at room temperature.
Storage Conditions	Store at room temperature.

**Specifications**

pH=2.25 +/- 0.2 (at 25°C)  
Conductivity=27 +/- 5 mS/cm

**Limited Product Warranty**

This warranty limits our liability to replacement of this product. No other warranties of any kind, express or implied, including, without limitation, implied warranties of merchantability or fitness for a particular purpose, are provided by Agilent. Agilent shall have no liability for any direct, indirect, consequential, or incidental damages arising out of the use, the results of use, or the inability to use this product.

**For Research Use Only. Not for use in diagnostic procedures. This certificate is a declaration of analysis at the time of manufacture.**

**Made in USA**

Agilent Technologies  
1834 State Highway 71 West  
Cedar Creek, TX 78612 USA

For support within the US or Canada: 1-800-227-9770-3-4  
For global support: [www.agilent.com/genomics/contactus](http://www.agilent.com/genomics/contactus)