



5188-5279 100758147



Agilent Technologies

Lot 100758147

Page 1 of 1

Catalog Number	5188-5279
Product Name	Two-Color RNA Spike-In Kit
Expiration Date	2017-05-30
Quantity	1 Kit
Certified By	Tim Sanchez
Quality Controlled By	Matt Huffman
Shipping Conditions	Shipped on dry ice.
Storage Conditions	Store at - 80°C upon receipt.

Materials Provided

Spike A Mix, 10 µl
Spike B Mix, 10 µl
Dilution Buffer, 1.2 ml

Description

The Agilent RNA Spike-In Kit was developed to provide positive controls for monitoring the microarray workflow from sample amplification and labeling to microarray processing.

Test Conditions

RNA samples are spiked using the Two-Color RNA Spike-In Kit and then labeled with two dyes. The labeled cRNA is hybridized to Gene Expression microarrays to functionally verify the Spike-In reagents.

Limited Product Warranty

This warranty limits our liability to replacement of this product. No other warranties of any kind, express or implied, including, without limitation, implied warranties of merchantability or fitness for a particular purpose, are provided by Agilent. Agilent shall have no liability for any direct, indirect, consequential, or incidental damages arising out of the use, the results of use, or the inability to use this product.

For Research Use Only. Not for use in diagnostic procedures. This certificate is a declaration of analysis at the time of manufacture.

Made in USA

Agilent Technologies
1834 State Highway 71 West
Cedar Creek, TX 78612 USA

For support within the US or Canada: 1-800-227-9770-3-4
For global support: www.agilent.com/genomics/contactus