



5188-6410 101187061

**Agilent Technologies**

Lot 101187061

Page 1 of 1

---

|                       |                                                     |
|-----------------------|-----------------------------------------------------|
| Catalog Number        | 5188-6410                                           |
| Product Name          | Multi Affinity Removal Column Human-7, 4.6 x 100 mm |
| Expiration Date       | 2022-08-30                                          |
| Quantity              | 1 Column, 1.66 ml resin bed                         |
| Certified By          | Kelly Flesher                                       |
| Quality Controlled By | Matt Huffman                                        |
| Shipping Conditions   | Shipped with ice pack.                              |
| Storage Conditions    | Store at 2–8°C upon receipt.                        |
| Serial Number         | 101187061B                                          |
| Hu-7 Resin Lot Number | 0006547086                                          |
| Date Packed           | 2020-08-14                                          |

**Description**

Multiple Affinity Removal Column Human-7 is packed with affinity ligand-modified resin that simultaneously removes human albumin, IgG,  $\alpha$ 1-antitrypsin, IgA, transferrin, haptoglobin, and fibrinogen from human biological samples, such as serum, plasma, and cerebrospinal fluid.

Each column has a unique serial number that indicates when the column was packed, and the lot number of Human-7 (Hu-7) resin listed on this certificate indicates the lot of Human-7 resin that has been packed into the column.

**Column Specifications**Dynamic Binding Capacity of Packing Media

| Column Size  | Capacity for Human Plasma* |
|--------------|----------------------------|
| 4.6 x 100 mm | $\leq 70 \mu\text{l}$      |

\*Determined using human plasma obtained from Rockland Immunologicals, Inc. (product number: D519-04-0500)

Physical Inspection

The finished columns are visually inspected to verify correct labeling and assembly.

**Limited Product Warranty**

This warranty limits our liability to replacement of this product. No other warranties of any kind, express or implied, including, without limitation, implied warranties of merchantability or fitness for a particular purpose, are provided by Agilent. Agilent shall have no liability for any direct, indirect, consequential, or incidental damages arising out of the use, the results of use, or the inability to use this product.

**For Research Use Only. Not for use in diagnostic procedures. This certificate is a declaration of analysis at the time of manufacture.**

**Made in USA**

Agilent Technologies  
1834 State Highway 71 West  
Cedar Creek, TX 78612 USA

For support within the US or Canada: 1-800-227-9770-3-4  
For global support: [www.agilent.com/genomics/contactus](http://www.agilent.com/genomics/contactus)