



5188-8826 0006713393



Lot 0006713393

Page 1 of 1

Catalog Number 5188-8826
Product Name Multi Affinity Removal Column HSA/IgG, 4.6 x 50 mm
Expiration Date 2025-01-30
Quantity 1 Column
Certified By Kelly Flesher
Quality Controlled By Matt Huffman
Shipping Conditions Shipped with ice pack.
Storage Conditions Store at 2–8°C upon receipt.
Serial Number 101523011N
HSA/IgG Resin Lot Number 0006660146
Date Packed 2023-01-26

Description

Multi Affinity Removal Column HSA/IgG is packed with affinity ligand-modified resin that removes human albumin and IgG from human biological samples, such as serum, plasma, and cerebrospinal fluid.

Each column has a unique serial number that indicates when the column was packed, and the lot number of HSA/IgG resin listed on this certificate indicates the lot of HSA/IgG resin that has been packed into the column.

Column Specifications

Dynamic Binding Capacity of Packing Media

Column Size	Capacity for Human Serum*
4.6 x 50 mm	≤ 100 µl

*Determined using pooled human serum obtained from Sigma (product number: S-7023)

Physical Inspection

The finished columns are visually inspected to verify correct labeling and assembly.

Limited Product Warranty

This warranty limits our liability to replacement of this product. No other warranties of any kind, express or implied, including, without limitation, implied warranties of merchantability or fitness for a particular purpose, are provided by Agilent. Agilent shall have no liability for any direct, indirect, consequential, or incidental damages arising out of the use, the results of use, or the inability to use this product.

For Research Use Only. Not for use in diagnostic procedures. This certificate is a declaration of analysis at the time of manufacture.

Made in USA

Agilent Technologies
1834 State Highway 71 West
Cedar Creek, TX 78612 USA

For support within the US or Canada: 1-800-227-9770-3-4
For global support: www.agilent.com/genomics/contactus