



5190-4661 0006905135



Lot 0006905135

Page 1 of 1

Catalog Number	5190-4661
Product Name	SureSelect XT Human Methyl Seq Library
Expiration Date	2031-05-31
Quantity	16 Reactions
Certified By	Meagan Stevens
Quality Controlled By	Zulema Bates
Shipping Conditions	Shipped on dry ice.
Storage Conditions	Store at -80°C upon receipt.

Description

The SureSelect XT Human Methyl-Seq Library is used in a DNA capture protocol with a SureSelect XT Methyl-Seq Reagent Kit to prepare a DNA methylation sequencing library for Illumina sequencing platforms.

Test Conditions

The RNA probe library integrity is verified using an Agilent TapeStation system or Agilent 2100 Bioanalyzer system. The RNA probe library is functionally tested in a SureSelect DNA capture experiment with performance verified by sequencing.

Limited Product Warranty

This warranty limits our liability to replacement of this product. No other warranties of any kind, express or implied, including, without limitation, implied warranties of merchantability or fitness for a particular purpose, are provided by Agilent. Agilent shall have no liability for any direct, indirect, consequential, or incidental damages arising out of the use, the results of use, or the inability to use this product.

For Research Use Only. Not for use in diagnostic procedures. This certificate is a declaration of analysis at the time of manufacture.

Made in USA

Agilent Technologies
1834 State Highway 71 West
Cedar Creek, TX 78612 USA

For support within the US or Canada: 1-800-227-9770-3-4
For global support: www.agilent.com/genomics/contactus