



5190-8046 0006299820



Agilent Technologies

Lot 0006299820

Page 1 of 1

Catalog Number	5190-8046
Product Name	HaloPlex 0.5-2.5 Mb OR <0.5 Mb with >20,000 probes, ILM, Box 1, 48 Reactions
Expiration Date	2017-04-30
Certifications	Certified by Tim Sanchez. Quality Controlled by Matt Huffman.
Shipping/Storage Conditions	Shipped on dry ice. Storage Temperature: -20°C.
Design ID	16592-1444633942
PCR Cycle Recommendation	Run 21 PCR cycles during the HaloPlex protocol amplification step for the probe included in this shipment.

Hyb Time Recommendation

This custom HaloPlex design contains >20,000 probes and requires a hybridization time of 16 hours.

Test Conditions

The target enrichment kit undergoes a series of functional tests including analytical restriction enzyme digestion and enriched DNA imaging to verify that the kit meets the required quality specifications.

Limited Product Warranty

This warranty limits our liability to replacement of this product. No other warranties of any kind, express or implied, including, without limitation, implied warranties of merchantability or fitness for a particular purpose, are provided by Agilent. Agilent shall have no liability for any direct, indirect, consequential, or incidental damages arising out of the use, the results of use, or the inability to use this product.

Laboratory Reagent. Not for use in diagnostic procedures. This certificate is a declaration of analysis at the time of manufacture.

Made in USA

Agilent Technologies
1834 State Highway 71 West
Cedar Creek, TX 78612 USA

For support within the US or Canada: 1-800-227-9770-3-4
For global support: www.agilent.com/genomics/contactus