



5190-6455 0006172383



Agilent Technologies

Lot 0006172383

Page 1 of 1

Catalog Number	5190-6455
Product Name	SureSelect XT2 Human All Exon v5 + MHC
Expiration Date	2016-12-30
Quantity	16 Reactions
Certified By	Todd Parsons
Quality Controlled By	Tricia Molina
Shipping Conditions	Shipped on dry ice.
Storage Conditions	Store at -80°C upon receipt.

Description

The SureSelect Target Enrichment System is a DNA capture technology developed to isolate genomic sequences of interest for downstream analysis by next-generation sequencing (NGS).

Test Conditions

Exon v5 base library RNA is validated by sequencing. Synthesis and capture quality of additional MHC RNA content are validated by QPCR.

Limited Product Warranty

This warranty limits our liability to replacement of this product. No other warranties of any kind, express or implied, including, without limitation, implied warranties of merchantability or fitness for a particular purpose, are provided by Agilent. Agilent shall have no liability for any direct, indirect, consequential, or incidental damages arising out of the use, the results of use, or the inability to use this product.

For Research Use Only. Not for use in diagnostic procedures. This certificate is a declaration of analysis at the time of manufacture.

Made in USA

Agilent Technologies
1834 State Highway 71 West
Cedar Creek, TX 78612 USA

For support within the US or Canada: 1-800-227-9770-3-4
For global support: www.agilent.com/genomics/contactus