



Lot 0006302728

Page 1 of 1

Catalog Number	5190-7909
Product Name	Brilliant HRM Ultra-Fast Loci Master Mix, Sample size
Expiration Date	2017-01-30
Quantity	15 Reactions (150 µl of a 2× master mix)
Certified By	Dana Faver
Quality Controlled By	Matt Huffman
Shipping Conditions	Shipped on dry ice.

Description

The Brilliant HRM Ultra-Fast Loci Master Mix is for real-time PCR experiments that use high-resolution melt (HRM) analysis to distinguish between alleles at a SNP locus.

Storage Conditions

Store at -20°C upon receipt. After thawing, the master mix may be stored at 4°C for up to one month or returned to -20°C for long term storage. The master mix is light sensitive and should be kept away from light whenever possible.

Test Conditions

The 2× master mix is tested in an Allele Discrimination experiment with a class IV SNP on the AriaMx Real-Time PCR System. The melt curves are visually analyzed for clear separation of the homoalleles.

Contamination Test Conditions

The 2× master mix is tested extensively to confirm the absence of DNA contamination.

Notice To Purchaser

This product is provided under an agreement between Biotium, Inc. and Agilent Technologies, Inc. and the manufacture, use, sale or import of this product is the subject to one or more of pending patent applications owned by Biotium and Allelogic. The purchase of this product conveys to the buyer the non-transferable right to use the purchased amount of the product and components of the product in research conducted by the buyer, where such research does not include testing, analysis or screening services for any third party in return for compensation on a per test basis. The buyer cannot sell or otherwise transfer (a) this product (b) its components or (c) or otherwise use this product or its components or materials made using this product or its components for Commercial Purposes. Commercial Purposes means any activity by a Party for consideration and may include, but is not limited to: (1) use of the product or its components in manufacturing; (2) use of the product or its components to provide a service, information, or data; (3) use of the product or its components for therapeutic, diagnostic or prophylactic purposes; or (4) resale of the product or its components, whether or not such product or its components are resold for use in research. For information on purchasing a license to this product for purposes other than research, contact Biotium, Inc., Business Development, 3423 Investment Blvd, Suite 8, Hayward, CA 94545 Tel: 510-265-1027, Fax: 510-265-1352.

Limited Product Warranty

This warranty limits our liability to replacement of this product. No other warranties of any kind, express or implied, including, without limitation, implied warranties of merchantability or fitness for a particular purpose, are provided by Agilent. Agilent shall have no liability for any direct, indirect, consequential, or incidental damages arising out of the use, the results of use, or the inability to use this product.

For Research Use Only. Not for use in diagnostic procedures. This certificate is a declaration of analysis at the time of manufacture.

Made in USA

Agilent Technologies
1834 State Highway 71 West
Cedar Creek, TX 78612 USA

For support within the US or Canada: 1-800-227-9770-3-4
For global support: www.agilent.com/genomics/contactus