



5190-8012 0006347137



Agilent Technologies

Lot 0006347137

Page 1 of 1

Catalog Number	5190-8012
Product Name	ClearSeq Comprehensive Cancer XT
Expiration Date	2020-03-30
Quantity	96 Reactions
Certified By	Todd Parsons
Quality Controlled By	Matt Huffman
Shipping Conditions	Shipped on dry ice.
Storage Conditions	Store at -80°C upon receipt.

Description

This 0.79 Mb ClearSeq capture library is used with the SureSelect XT or SureSelect QXT Target Enrichment System to isolate genomic sequences of interest for downstream analysis by next-generation sequencing (NGS).

Test Conditions

Library RNA integrity and concentration is verified using the Agilent 2100 Bioanalyzer. The RNA library is functionally tested in a SureSelect DNA capture experiment, with performance verified by sequencing.

Limited Product Warranty

This warranty limits our liability to replacement of this product. No other warranties of any kind, express or implied, including, without limitation, implied warranties of merchantability or fitness for a particular purpose, are provided by Agilent. Agilent shall have no liability for any direct, indirect, consequential, or incidental damages arising out of the use, the results of use, or the inability to use this product.

For Research Use Only. Not for use in diagnostic procedures. This certificate is a declaration of analysis at the time of manufacture.

Made in USA

Agilent Technologies
1834 State Highway 71 West
Cedar Creek, TX 78612 USA

For support within the US or Canada: 1-800-227-9770-3-4
For global support: www.agilent.com/genomics/contactus