



G111401 0006616894



Lot **0006616894**
Page 1 of 1

Catalog Number	G111401
Product Name	SureFISH ROS1 BA P20
Materials Provided	G111401-85510, 1 tube (20 µl)
Description	ROS1 Red: genomic coordinates: Human Chr6:116510643-117609677 ROS1 Green: genomic coordinates: Human Chr6:117747013-118899513
Manufacturing Date	2021-08-30
Expiration Date	2023-08-30
Certified By	Todd Parsons
Quality Controlled By	Matt Huffman
Test Conditions	The translocation probes are tested by FISH to ensure detection of and specificity for the stated genomic coordinates in both interphase nuclei and metaphase spreads. For a representative image of FISH results obtained with this probe sequences, see the SureFISH website at www.agilent.com/genomics/surefish

Made in USA
Agilent Technologies
1834 State Highway 71 West
Cedar Creek, TX 78612 USA

For support within the US or Canada: 1-800-227-9770-3-4-2
For global support: www.agilent.com/genomics/contactus

013222-SureFISH-Transloc

Volume	20 µl
Shipping Conditions	Shipped on dry ice.
Storage Conditions	Store at -25°C to -15°C upon receipt. Protect from light whenever possible.
Warning	Fluorophores are light-sensitive and will readily photobleach. Handle all solutions and slides containing labeled probes under reduced light. Perform all incubations and washes in the dark whenever possible. Ensure that all slides are light protected prior to imaging.
Notice to Purchaser	Patent information for this product is available on the web at www.agilent.com/genomics/aboutsurefish .
Limited Product Warranty	This warranty limits our liability to replacement of this product. No other warranties of any kind, express or implied, including, without limitation, implied warranties of merchantability or fitness for a particular purpose, are provided by Agilent. Agilent shall have no liability for any direct, indirect, consequential, or incidental damages arising out of the use, the results of use, or the inability to use this product.

Analyte Specific Reagent. Analytical and performance characteristics are not established.
This certificate is a declaration of analysis at the time of manufacture.