



G111600-8 0006704300



Lot **0006704300**  
Page 1 of 1

---

<b>Catalog Number</b>	<b>G111600-8</b>
<b>Product Name</b>	ALK IQFISH Break-Apart Probe
<b>Materials Provided</b>	G111600-85510, 1 tube (200 µl)
<b>Description</b>	ALK 3' Orange-Red, genomic coordinates: Human Chr2:29146786-29446528 ALK 5' Green, genomic coordinates: Human Chr2:29446949-30045655
<b>Manufacturing Date</b>	2023-05-30
<b>Expiration Date</b>	2025-05-30
<b>Certified By</b>	Todd Parsons
<b>Quality Controlled By</b>	Matt Huffman
<b>Test Conditions</b>	The probe has been tested by FISH and viewed at a magnification of 600× to ensure detection of and specificity for the stated genomic coordinates in interphase nuclei, metaphase spreads, and slices of FFPE-embedded tissue. For a representative image of FISH results obtained with this probe sequences, see the SureFISH website at <a href="http://www.agilent.com/genomics/surefish">www.agilent.com/genomics/surefish</a>  For support within the US or Canada: 1-800-227-9770-3-4-2 For global support: <a href="http://www.agilent.com/genomics/contactus">www.agilent.com/genomics/contactus</a>

<b>Shipping Conditions</b>	Shipped on dry ice.
<b>Storage Conditions</b>	Store at -25°C to -15°C upon receipt. Protect from light whenever possible.
<b>Warning</b>	Fluorophores are light-sensitive and will readily photobleach. Handle all solutions and slides containing labeled probes under reduced light. Perform all incubations and washes in the dark whenever possible. Ensure that all slides are light protected prior to imaging.
<b>Notice to Purchaser</b>	Patent information for this product is available on the web at <a href="http://www.agilent.com/genomics/aboutsurefish">www.agilent.com/genomics/aboutsurefish</a> .
<b>Limited Product Warranty</b>	This warranty limits our liability to replacement of this product. No other warranties of any kind, express or implied, including, without limitation, implied warranties of merchantability or fitness for a particular purpose, are provided by Agilent. Agilent shall have no liability for any direct, indirect, consequential, or incidental damages arising out of the use, the results of use, or the inability to use this product.



**Made in USA**  
Agilent Technologies, Inc.  
5301 Stevens Creek Blvd, Santa Clara, CA 95051, USA  
Manufactured at:  
1834 State Hwy 71 W, Cedar Creek, TX 78612, USA  
[www.agilent.com](http://www.agilent.com)



This certificate is a declaration of analysis at the time of manufacture.