



G111601-8 0006269403



**Agilent Technologies**

Lot **0006269403**

Page 1 of 1

---

<b>Catalog Number</b>	<b>G111601-8</b>
<b>Product Name</b>	ROS1 IQFISH Break-Apart Probe
<b>Materials Provided</b>	G111601-85510, 1 tube (200 µl)
<b>Description</b>	ROS1 Orange-Red, genomic coordinates: Human Chr6:116510643-117609677 ROS1 Green, genomic coordinates: Human Chr6:117747013-118899513
<b>Manufacturing Date</b>	2015/04/30
<b>Expiration Date</b>	2017/04/30
<b>Certified By</b>	Todd Parsons
<b>Quality Controlled By</b>	Tricia Molina
<b>Test Conditions</b>	The probe has been tested by FISH and viewed at a magnification of 600× to ensure detection of and specificity for the stated genomic coordinates in interphase nuclei, metaphase spreads, and slices of FFPE-embedded tissue. For a representative image of FISH results obtained with this probe sequences, see the SureFISH website at <a href="http://www.agilent.com/genomics/surefish">www.agilent.com/genomics/surefish</a>

Agilent Technologies  
1834 State Highway 71 West  
Cedar Creek, TX 78612 USA

For support within the US or Canada: 1-800-227-9770-3-4-2  
For global support: [www.agilent.com/genomics/contactus](http://www.agilent.com/genomics/contactus)

**Shipping Conditions**

Shipped on dry ice.

**Storage Conditions**

Store at -25°C to -15°C upon receipt. Protect from light whenever possible.

**Warning**

Fluorophores are light-sensitive and will readily photobleach. Handle all solutions and slides containing labeled probes under reduced light. Perform all incubations and washes in the dark whenever possible. Ensure that all slides are light protected prior to imaging.

**Notice to Purchaser**

Patent information for this product is available on the web at [www.agilent.com/genomics/aboutsurefish](http://www.agilent.com/genomics/aboutsurefish).

**Limited Product Warranty**

This warranty limits our liability to replacement of this product. No other warranties of any kind, express or implied, including, without limitation, implied warranties of merchantability or fitness for a particular purpose, are provided by Agilent. Agilent shall have no liability for any direct, indirect, consequential, or incidental damages arising out of the use, the results of use, or the inability to use this product.



This certificate is a declaration of analysis at the time of manufacture.