G111603-8 0006879627

Agilent

Lot	0006879627

Catalog Number	G111603-8
Product Name	MET IQFISH Probe with CEP7
Materials Provided	G111603-85510, 1 tube (200 µl)
Description	MET Orange-Red, genomic coordinates: Human Chr:7:116192143-116591613 CEP7 Green, genomic coordinates: Human Chr7:66804212-67571167
Manufacturing Date	2026-02-26
Expiration Date	2029-02-28
Certified By	Todd Parsons
Quality Controlled By	Zulema Bates
Test Conditions	The probe has been tested by FISH and viewed at a magnification of 600× to ensure detection of and specificity for the stated genomic coordinates in interphase nuclei, metaphase spreads, and slices of FFPE-embedded tissue. For a representative image of FISH results obtained with this probe sequences, see the SureFISH website at www.agilent.com/genomics/surefish For support within the US or Canada: 1-800-227-9770-3-4-2 For global support: www.agilent.com/genomics/contactus
Shipping Conditions	Shipped on dry ice.
Storage Conditions	Store at -25ºC to -15ºC upon receipt. Protect from light whenever possible.
Warning	Fluorophores are light-sensitive and will readily photobleach. Handle all solutions and slides containing labeled probes under reduced light. Perform all incubations and washes in the dark whenever possible. Ensure that all slides are light protected prior to imaging.
Notice to Purchaser	Patent information for this product is available on the web at www.agilent.com/genomics/aboutsurefish.
Limited Product Warranty	This warranty limits our liability to replacement of this product. No other warranties of any kind, express or implied, including, without limitation, implied warranties of merchantability or fitness for a particular purpose, are provided by Agilent. Agilent shall have no liability for any direct, indirect, consequential, or incidental damages arising out of the use, the results of use, or the inability to use this product.

Made in USA
Agilent Technologies, Inc.
5301 Stevens Creek Blvd, Santa Clara, CA 95051, USA
Manufactured at:
1834 State Hwy 71 W, Cedar Creek, TX 78612, USA
www.agilent.com

CE IVD

This certificate is a declaration of analysis at the time of manufacture.