



Lot 0006297740

Page 1 of 1

Catalog Number G9414A
 Product Name IQFISH Fast Hybridization Buffer 900
 Expiration Date 2016-10-30
 Quantity 900 µl
 Certified By Todd Parsons
 Quality Controlled By Tricia Molina
 Shipping Conditions Shipped on dry ice.
 Storage Conditions Store at -25°C to -15°C upon receipt.

Description

For use in fast (1–2 hour) hybridization time fluorescence *in situ* hybridization (FISH) protocols.

Test Conditions

The buffer is tested in a FISH protocol to verify that the signal to noise ratio meets specifications.

Symbol Table			
	Catalog Number		Batch Code or Lot Number
	Temperature limitation		Use by date YYYY-MM-DD or YYYY-MM
	Manufacturer Address		Irritant

Limited Product Warranty

This warranty limits our liability to replacement of this product. No other warranties of any kind, express or implied, including, without limitation, implied warranties of merchantability or fitness for a particular purpose, are provided by Agilent. Agilent shall have no liability for any direct, indirect, consequential, or incidental damages arising out of the use, the results of use, or the inability to use this product.

For Laboratory Use. This certificate is a declaration of analysis at the time of manufacture.

Made in USA

Agilent Technologies
1834 State Highway 71 West
Cedar Creek, TX 78612 USA

For support within the US or Canada: 1-800-227-9770-3-4
For global support: www.agilent.com/genomics/contactus