



Lot 0006441029

Page 1 of 1

Catalog Number G9411A  
Product Name ISH Pepsin Kit  
Expiration Date 2020-03-30  
Quantity 48 ml Pepsin (2 mg/ml), 48 ml Pepsin Diluent (10X)  
Certified By Todd Parsons  
Quality Controlled By Matt Huffman  
Shipping Conditions Shipped on dry ice.  
Storage Conditions Store at -25°C to -15°C for duration of expiry date or at 2°C to 8°C for 60 days post receipt.

### Description

For use in *in situ* hybridization protocols.

### Test Conditions

The Pepsin and 10× Pepsin Diluent are functionally tested in a fluorescence in situ hybridization (FISH) assay to verify that the reagents meet performance specifications. The pH and appearance of the reagents are also evaluated to ensure they meet specifications.

| Symbol Table                                                                        |                        |                                                                                      |                                      |
|-------------------------------------------------------------------------------------|------------------------|--------------------------------------------------------------------------------------|--------------------------------------|
|  | Catalog Number         |  | Batch Code or Lot Number             |
|  | Temperature limitation |  | Use by date<br>YYYY-MM-DD or YYYY-MM |
|  | Manufacturer Address   |  | Flammable                            |
|  | Irritant               |  | Corrosive                            |

### Limited Product Warranty

This warranty limits our liability to replacement of this product. No other warranties of any kind, express or implied, including, without limitation, implied warranties of merchantability or fitness for a particular purpose, are provided by Agilent. Agilent shall have no liability for any direct, indirect, consequential, or incidental damages arising out of the use, the results of use, or the inability to use this product.

For Laboratory Use. This certificate is a declaration of analysis at the time of manufacture.

### Made in USA

Agilent Technologies  
1834 State Highway 71 West  
Cedar Creek, TX 78612 USA

For support within the US or Canada: 1-800-227-9770-3-4  
For global support: [www.agilent.com/genomics/contactus](http://www.agilent.com/genomics/contactus)