



G9411A 0006568404



Lot 0006568404

Page 1 of 1

Catalog Number	G9411A
Product Name	ISH Pepsin Kit
Expiration Date	2022-01-30
Quantity	48 ml Pepsin (2 mg/ml), 48 ml Pepsin Diluent (10X)
Certified By	Todd Parsons
Quality Controlled By	Matt Huffman
Shipping Conditions	Shipped on dry ice.
Storage Conditions	Store at -25°C to -15°C for duration of expiry date or at 2°C to 8°C for 60 days post receipt.

Description

For use in *in situ* hybridization protocols.

Test Conditions

The Pepsin and 10× Pepsin Diluent are functionally tested in a fluorescence in situ hybridization (FISH) assay to verify that the reagents meet performance specifications. The pH and appearance of the reagents are also evaluated to ensure they meet specifications.

Symbol Table			
	Catalog Number		Batch Code or Lot Number
	Temperature limitation		Use by date YYYY-MM-DD or YYYY-MM
	Manufacturer Address		Flammable
	Irritant		Corrosive

Limited Product Warranty

This warranty limits our liability to replacement of this product. No other warranties of any kind, express or implied, including, without limitation, implied warranties of merchantability or fitness for a particular purpose, are provided by Agilent. Agilent shall have no liability for any direct, indirect, consequential, or incidental damages arising out of the use, the results of use, or the inability to use this product.

For Laboratory Use. This certificate is a declaration of analysis at the time of manufacture.

Made in USA

Agilent Technologies
1834 State Highway 71 West
Cedar Creek, TX 78612 USA

For support within the US or Canada: 1-800-227-9770-3-4
For global support: www.agilent.com/genomics/contactus