

GPC Calibration Solution

Product Number: CLP-341

Page: 1 of 1

Lot Number: CF-4110

Lot Issue Date: Oct-2009

Expiration Date: Nov-2012

This Certified Reference Material (RM) was manufactured and verified in accordance with ULTRA's ISO 9001 registered quality system, and the analyte concentrations were verified by our ISO 17025 accredited laboratory. The true value and uncertainty value at the 95% confidence level for each analyte, determined gravimetrically, is listed below.

Analyte	CAS#	Analyte Lot	True Value
corn oil	008001-30-7	RM00924	300.5 ± 1.5 mg/mL
bis(2-ethylhexyl) phthalate	000117-81-7	D-23230	15.0 ± 0.1 mg/mL
pentachlorophenol	000087-86-5	07119HO	1.40 ± 0.01 mg/mL
perylene	000198-55-0	04215DY	0.100 ± 0.001 mg/mL
sulfur (elemental)	007704-34-9	04911TB	0.500 ± 0.003 mg/mL

Matrix: methylene chloride (dichloromethane)


ULTRA uses balances calibrated with weights traceable to NIST in compliance with ANSI/NCCL Z-540-1 and ISO 9001, and calibrated Class A glassware in the manufacturing of these standards.



ISO 17025
Accredited
A2LA
Cert. No. 0851-01

ISO 9001:2000
Registered
TUV USA, Inc.
Cert. No. 06-1004

250 Smith Street, North Kingstown, RI 02852 USA
401-294-9400 Fax: 295-2330
www.ultrasci.com



William J. Leary
Quality Assurance Manager