



5067-5584



Declaration of Conformity to Manufacturing Specification

Agilent part number:

5067-5584

Product name:

High Sensitivity D1000 ScreenTape

Storage conditions:

2 °C to 8 °C

Shipping conditions:

2 °C to 8 °C

Lot number:

0203148

Expiration date:

2024-10-15

Release date:

2024-01-13

Approved by:

Sebastian Pallmann

Sebastian Pallmann
Senior Quality Engineer
Agilent Technologies
Deutschland GmbH

Date:

2024-01-13

This product was manufactured and verified in accordance with an ISO 9001 certified quality management system and tested in accordance to ISO 2859-1 as applicable. The manufacture, inspection, and testing were performed by trained personnel. The equipment and materials used in the manufacturing process are identifiable and traceable. All associated test equipment is maintained and calibrated.

The following product characteristics were tested for conformity to manufacturing specifications.

Product characteristic	Method	Result
ScreenTape integrity	Visual inspection	Pass
Fragment migration behavior	Electrophoretical with DNA reference ladder	Pass
Sizing accuracy	Electrophoretical with DNA reference ladder	Pass
Sizing precision	Electrophoretical with DNA reference ladder	Pass
Resolution	Electrophoretical with DNA reference ladder	Pass

For Research Use Only. Not for use in diagnostic procedures.

Agilent Technologies
Deutschland GmbH
Hewlett-Packard-Str. 8
76337 Waldbronn, Germany

Made in Germany

For global support: www.agilent.com/en/contact-us/page

Template Number: 5067-5584 - DOC MS
Template Revision: B.00